

RILEVATORI DI TEMPERATURA INFRAROSSI



Rilevatori di temperatura
Funzionali alla protezione COVID-19



COD TEK 2006

Rilevatore di temperatura ad infrarossi

Modello No.

Termometro a infrarossi DIKANG-HG01-V1

Dimensioni

149 millimetri * 77 millimetri * 43 millimetri

Peso

Ca. 95g (escluda 2 batterie AAA)

Display

LCD

Gamma di Temperatura del corpo

32.0 °C - 42.9 °C

Range di Temperatura

0 °C - 100 °C

Memoria 32 misurazioni

Tempo di risposta < 1 secondo

Ambiente di lavoro

Temperatura ambiente: 16 °C - 35 °C; Umidità

Relativa: ≤ 85%

Trasporto e Condizioni Di Stoccaggio

-20 °C ~ 55 °C

Precisione

34.0-34.9 °C: ± 0.3 °C; 34.0-42.0 °C: ± 0.2 °C;
42.1-42.9 °C: ± 0.3 °C

Risoluzione 0.1 °C

Tensione 2x1.5 V; Batteria formato AAA

Spegnimento automatico ≤ 15 secondi

Imballaggio 1 pc in una scatola di colore

CE认证发证机构为 ECM

Certificate of Compliance

No. 01200404/HIED066

Certificate's Holder: Hunan Honggao Electronic Technology Co., Ltd.
Block 3, Comprehensive Industrial Park, Tenghui Pioneer Park, Nanxin Economic Development Zone, Yiyang, Hunan Province.

Certification ECM Mark:



Product: Infrared Forehead Thermometer
Model(s): HG01, HG02, HG03, HG04, HG05, HG06, HG01, HG02, HG03, HG04, HG05, HG06, F101, F102, F103

Verification to: Standard: EN 60601-1:2006+A11:2011+A12:2014, EN 55011:2014+A1:2017, EN 60601-1-2:2017, EN 61000-3-2:2014, EN 61000-3-3:2013
related to CE Directive(s): 2014/53/EU (Low Voltage), 2014/30/EU (Electromagnetic Compatibility)

Remarks: This document has been issued on a voluntary basis and upon request of the manufacturer. It is for information that the technical documentation received from the manufacturer is satisfactory for the requirements of the ECM Certification Mark. The conformity mark above can be affixed on the products accordingly to the ECM regulation about to review and to use.

Additional Information and clarification about the Marking:

The manufacturer is responsible for the CE Marking process, and if necessary, must refer to a notified body. This document has been issued on the basis of the regulation on ECM Voluntary Mark for the certification of products. PG01_ECM rev.3 available at: www.ecmcert.it

Issuance date: 04 April 2020

Expiry date: 03 April 2025

Reviewer: Technical expert Amanda Payne
Approver: ECM Service Director Luca Sestini
Via C.so. 243 - Loc. Castello di Senovate - 40030 Valsamoggia (BO) - Italy
Tel: +39 051 6700141 - Fax: +39 051 6700156 - Email: info@ecmcert.it - www.ecmcert.it

Registrazione Ministero
della Salute n. 1964590



TEKNOLEDIMPIANTI SRLS
Soc Unipersonale
Via Delle Vigne di Morena 90 00118 Roma
P.IVA 14035911008 Cod Univ M5UXCR1
Web Site : WWW.TEKNOLEDIMPIANTI.IT
e-mail : info@teknoleimpianti.it
Pec : tteknoleimpianti@legalmail.it